OB. NR	LOC.	

MANUFACTURER'S CERTIFICATE OF CONFORMANCE

-2238--001-

2684502

We certify that Brüel & Kjær

Serial No

has been tested and passed all production tests, confirming compliance with the manufacturer's published specification at the date of the test.

The final test has been performed using calibrated equipment, traceable to National or International Standards or by ratio measurements.

Brüel & Kjær is certified under ISO 9001:2000 assuring that all calibration data for test equipment are retained on file and are available for inspection upon request. 07-sep-2009

Nærum

Torben Bjørn Vice President Operations

Please note that this document is not a calibration certificate, for information on our calibration services please contact your nearest Brüel & Kjær Service Center.

BA 0238 - 15

WORLD HEADQUARTERS: DK-2850 Nærum \cdot Denmark Telephone: +45 45 80 05 00 \cdot Fax: +45 45 80 14 05 \cdot http://www.bksv.com \cdot e-mail: info@bksv.dk



OB. NR	LOC	

2684503

MANUFACTURER'S CERTIFICATE OF CONFORMANCE

We certify that Brüel & Kjær

-2238--001-

Serial No

has been tested and passed all production tests, confirming compliance with the manufacturer's published specification at the date of the test.

The final test has been performed using calibrated equipment, traceable to National or International Standards or by ratio measurements.

Brüel & Kjær is certified under ISO 9001:2000 assuring that all calibration data for test equipment are retained on file and are available for inspection upon request. 07-sep-2009

Nærum

Torben Bjørø Vice President Operations

Please note that this document is not a calibration certificate, for information on our calibration services please contact your nearest Brüel & Kjær Service Center.

BA 0238 - 15

Certificate No.: C095419

Certificate of Calibration

This is to certify that the equipment

Description: Acoustical Calibrator

Manufacturer: Bruel & Kjaer

Model No.: 4231

Serial No.: 2385180

has been calibrated for the specific items and ranges. The results are shown in the Calibration Report No. C095419.

The equipment is supplied by

Co. Name: Atkins China Limited

Address: 5/F., Wharf T&T Centre, Harbour City, Tsim Sha Tsui, Kowloon

Date of Issue: 9 October 2009

Certified by:

K C Lee

The test equipment used for calibration are traceable to the National Standards as specified in this report. This report shall not be reproduced except in full and with prior written approval from this laboratory.

Tel: 2927 2606

Fax: 2744 8986

E-mail: callab@suncreation.com

Website: www.suncreation.com



輝創工程有限公司

Sun Creation Engineering Limited Calibration and Testing Laboratory

Report No.: C095419

Calibration Report

ITEM TESTED

DESCRIPTION

: Acoustical Calibrator

MANUFACTURER:

Bruel & Kjaer

MODEL NO.

4231

SERIAL NO.

: 2385180

TEST CONDITIONS

AMBIENT TEMPERATURE : $(23 \pm 2)^{\circ}$ C

RELATIVE HUMIDITY: $(55 \pm 20)\%$

LINE VOLTAGE

TEST SPECIFICATIONS

Calibration check

DATE OF TEST: 8 October 2009

JOB NO.: IC09-2609

TEST RESULTS

The results apply to the particular unit-under-test only.

All results are within manufacturer's specification.

The results are detailed in the subsequent page(s).

The test equipment used for calibration are traceable to National Standards via:

- The Government of The Hong Kong Special Administrative Region Standard & Calibration Laboratory
- Agilent Technologies, USA
- Fluke Everett Service Center, USA
- Rohde & Schwarz Laboratory, Germany

Tested by:

Date: 9 October 2009

The test equipment used for calibration are traceable to the National Standards as specified in this report. This report shall not be reproduced except in full and with prior written approval from this laboratory.



輝創工程有限公司

Sun Creation Engineering Limited Calibration and Testing Laboratory

Report No.: C095419

Calibration Report

- 1. The unit-under-test (UUT) was allowed to stabilize in the laboratory for over 24 hours before the commencement of the test.
- 2. The results presented are the mean of 3 measurements at each calibration point.
- 3. Test equipment:

Equipment ID CL130 CL281 TST150A

<u>Description</u>
Universal Counter
Multifunction Acoustic Calibrator

Measuring Amplifier

Certificate No. C093122 DC090052 C080751

4. Test procedure: MA100N.

5. Results:

5.1 Sound Level Measurement

UUT Nominal Value	Measured Value (dB)	Mfr's Spec. (dB)	Uncertainty of Measured Value (dB)
94 dB, 1 kHz	94.0	± 0.2	± 0.2
114 dB, 1 kHz	114.0		- 0.2

5.2 Frequency Accuracy

UUT Nominal Value (kHz)	Measured Value (kHz)	Mfr's Spec.	Uncertainty of Measured Value (Hz)
1	1.000 0	1 kHz ± 0.1 %	± 0.1

Remark: - The uncertainties are for a confidence probability of not less than 95 %.

Note:

The values given in this Calibration Report only relate to the values measured at the time of the test and any uncertainties quoted will not include allowance for the equipment long term drift, variations with environment changes, vibration and shock during transportation, overloading, mis-handling, or the capability of any other laboratory to repeat the measurement. Sun Creation Engineering Limited shall not be liable for any loss or damage resulting from the use of the equipment.

The test equipment used for calibration are traceable to the National Standards as specified in this report. This report shall not be reproduced except in full and with prior written approval from this laboratory.